

DECLARATION OF CONFORMITY TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES



MANUFACTURER:

Sino-Hero(Shenzhen) Bio-Medical Electronics Co., LTD
A area,Second Floor, First
Building, Tongkangfu Industrial Park, Yingrenshi
Community, Shiyao Sub-District, Bao'an District,
518108,Shenzhen P. R. China

MEDICAL DEVICE:

Pulse Oximeter/ S8

UMDNS CODE

17148

CLASSIFICATION - ANNEX IX:

CLASS IIb, RULE 10

CONFORMITY ASSESSMENT ROUTE:

ANNEX II EXCLUDING 4

WE, (**Sino-Hero(Shenzhen) Bio-Medical Electronics Co., LTD**), HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC CONCERNING MEDICAL DEVICES.

STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.

NOTIFIED BODY:

TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER

CE 0123

(EC) CERTIFICATE(S):

G1 071165 0013 Rev.02



EUROPEAN REPRESENTATIVE:

Shanghai International Holding Corp. GmbH(Europe)
Eiffestrasse 80, 20537 Hamburg, Germany


START OF CE-MARKING:

2009-11-19

PLACE, DATE OF DECLARATION:

SHENZHEN, 2019-11-04

SIGNATURE:


BEN LIN
VICE GENERAL MANAGER

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC CONCERNING
MEDICAL DEVICES**

All applicable harmonized Standard as bellows:

IEC 60601-1:2005 + A1: 2012

IEC 60601-1-2:2007

ISO 10993-1:2009

ISO 15223-1:2012

EN ISO 14971:2012

ISO 80601-2-61:2011